

CERTIFICATE OF CALIBRATION

Model: 1/2

Wrench Capacity: 28-210Nm **Inspector:** Andy Pan

Date of Calib: 2026/01/10 **Unit:** N • m

Set Point	(CW) ±4%		Actual Readings		
	Min	Max			
42	40.74	43.26	39.1	39.3	39.1
130	124.8	135.2	126.2	126.4	126.6
210	201.6	218.4	209.1	290.3	209.1

Testing Standard No: DIN-ISO-6789

Testing is in compliance with international standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: BE0002878

The uncertainty of the equipment at 20 °C is < ±3%

Inspector Approval:

